



# **Alachua County Budget and Fiscal Services Procurement**

Larry M. Sapp, CPPB  
Procurement Manager

Darryl R. Kight, CPPB  
Procurement Supervisor

May 21, 2019

RE: Addendum #3  
RFP 20-96 Annual Laboratory Analysis Services

Dear Sir/Madam:

Please be aware of the following clarifications regarding the above referenced Bid:

Q #1: Section 3.4.6 specifically lists that e-coli and fecal coliform must be conducted using membrane filter methodology. AEL plans to add NELAP certification to run these analyses by the IDEXX Quanti-Tray system by late next year. Is this a deviation that the county would consider accepting at that time?

A #1: At this time, the County's preference on e-coli and fecal coliform analysis is by membrane filter methodology. IDEXX Quanti-Tray system can be listed as an alternative method in Exhibit A 1-B.

The above **Bid Opening Date** has been extended to **June 19, 2019, at 2:00 pm**. All Bids must be received at the Procurement Department, 3<sup>rd</sup> Floor Administration Building, 12 SE 1<sup>st</sup> Street, Gainesville, FL 32601-6893.

MANDATORY PREBID has been scheduled for Tuesday, June 4, 2019 at 1:00 pm. Location is County Administration Building, 12 SE 1<sup>st</sup> Street, Third Floor Conference Room, Gainesville, Florida.

**Vendors who attended the MANDATORY PREBID May 1, 2019, do not have to attend the prebid scheduled for June 4, 2019.**

**NOTE:** You should acknowledge receipt of this addendum on your Bid Form.

**End of Addendum #3**

Sincerely,

Mandy Mullins  
Procurement Agent

MM